



Doc. nr. REA-S004-00-08-0511

Versie 2.0


Datum 06 May 2021 13:15:04 (GMT+01:00)

Pagina 1 van 3

Document eigenaar BBR

Titel BSA 22% release

K 1106



Analysis	Requirement	Result
Specificity	Negative with A and B cells and the cells of the Screening panel 123 tested with AB serum	Conform requirements
Titer	BSA 22% enhances reaction with rhesus antibodies significantly compared with saline method	Conform requirements
Growthtest	no growth	Conform

Date: 09 June 2021

Approved by QA/QC: [Signature]

VERTROUWELIJK