



Doc. nr. REA-S004-00-12-0515

Versie 2.0

Datum 06 May 2021 13:40:23 (GMT+01:00)

Pagina 1 van 2

Document eigenaar BBR

Titel PeliLISS release

K 1110



Analysis	Requirement	Result
Specificity	Negative with A and B cells and the cells of the Screening panel 123 x AB serum	Conform requirements
Titer	Reaction IAT with PeliLISS comparable or stronger than IAT with BSA 22%	Conform requirements
Growth test	no growth	Conform

Date

: 16 May 2022

Approved by QA/QC

:

VERTROUWELIJK