



Doc. nr. REA-S004-01-54-0565

Versie 2.0

Datum 06 May 2021 12:26:23 (GMT+02:00)

Pagina 1 van 2

Document eigenaar BBR


Titel Pelikloon anti-N (IgG) monoclonal release

K 1313

Clone number BO3



Analysis	Requirement	Result
Specificity	Negative with cells negative for the N antigen	Conform requirements
Avidity	Tube method	Conform requirements
	NN $\geq 2+$ MN $\geq 2+$	
Titer	Tube method	Conform requirements
	MN $\geq 1:4$	
Growthtest	no growth	Conform

Date : 06 December 2022 Approved by QA/QC : 

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 01 dec 2022 at: 09:46:42 am ; Printed by M.BOUWMAN