



Doc. nr. REA-S004-01-54-0565

Versie 2.0

Datum 06 May 2021 12:26:23 (GMT+02:00)

Pagina 1 van 2

Document eigenaar BBR

Titel Pelikloon anti-N (IgG) monoclonal release

K 1313

Clone number BO3

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

**Pelikloon anti-N (IgG) monoclonal** v.l.  
Clone: BO3 2 ml

IVD 2-8°C REF K1313 CE  
LOT 8000458973 2024-09-10

Analysis	Requirement	Result
Specificity	Negative with cells negative for the N antigen	Conform requirements
Avidity	Tube method	Conform requirements
	NN $\geq 2+$	
	MN $\geq 2+$	
Titer	Tube method	Conform requirements
	MN $\geq 1:4$	
Growthtest	no growth	Conform

Date

: 07 March 2023

Approved by QA/QC

:

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 01 mrt 2023 at: 08:26:30 am ; Printed by M.BOUWMAN