

**Titel Cellbind Screen release**

K 7000



Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+	Conform requirements
	Results with anti- Fy <sup>a</sup> , anti- Fy <sup>b</sup> and anti-s are at least comparable with reference method	Conform requirements

Date

: 16 May 2023

Approved by QA/QC

: 

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 15 feb 2023 at: 07:37:28 am ; Printed by M.BOUWMAN