


Titel **MASPAT kit release**

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

K 1360

MASPAT kit

Monoclonal Antibody Solid Phase  
Platelet Antibody Test



REF K1360

LOT 8000459475



2-8 °C

2024-08-21

(01)1872068203  
0025(10)800045  
9475(17)240821


Component	Requirement	Result
Positive control Lot: 8000458519	Positive reaction with: platelets Titration with: platelets $\geq$ 1:4	Conform requirements  Conform requirements
Negative control Lot: 8000458520	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000458517	Titration with: platelets $\geq$ 1:4	Conform requirements
Microplates Lot: 8000458104	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000458518	Conductivity 3.4 – 4.8 mScm <sup>-1</sup> pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date

: 26 May 2023

Approved by QA/QC

: 

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 04 mei 2023 at: 10:53:10 am ; Printed by M.BOUWMAN