

Titel **Pelikloon anti-k (IgM) monoclonal release**

K 1310

Clone nr. Lk1




Analysis	Requirement	Result
Specificity	Tube method Negative with cells negative for the k antigen	Conform requirements
Avidity	Tube method	
	kk $\geq 2+$ Kk $\geq 2+$	Conform requirements Conform requirements
Titer	Tube method	
	Kk $\geq 1:8$	Conform requirements
Growthtest	no growth	Conform

Date

: 03 July 2023

Approved by QA/QC

:



VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 06 juni 2023 at: 09:03:13 am ; Printed by M.BOUWMAN