



Doc. nr. REA-S004-01-49-0560

Versie 2.0

Datum 06 May 2021 11:56:07 (GMT+02:00)

Pagina 1 van 3

Document eigenaar BBR

Titel Pelikloon anti-K (IgM) monoclonal release

K 1199

Clone number MS56



Analysis	Requirement	Result
Specificity	Negative with cells negative for the K antigen	Conform requirements
Avidity	Kk	
	Tube method $\geq 2+$	Conform requirements
	Microplate method $\geq 1+$	Conform requirements
	Cellbind Screen method $\geq 2+$	Conform requirements
Titer	Cellbind Direct Type method $\geq 2+$	Conform requirements
	Kk	
	Tube method $\geq 1:16$	Conform requirements
	Microplate method $\geq 1:8$	Conform requirements
Growthtest	no growth	Conform
External release	released by Notified Body	Released Y/N

Date : 04 September 2023 Approved by QA/QC

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 14 Jul 2023 at: 11:36:12 am ; Printed by M.BOUWMAN