


**Titel Pelikloon anti-Cw (IgM) monoclonal release**

K1201

Clone nr. MS110



Analysis	Requirement	Result
Specificity	Tested in tube method, Cellbind Screen method and Cellbind Direct Type method  Negative with cells negative for the C <sup>w</sup> antigen	Conform requirements
Avidity	Tested in tube method, Cellbind Screen method and Cellbind Direct Type method  C <sup>w</sup> + ≥ 3+	Conform requirements
Titer	Tube method  C <sup>w</sup> + ≥ 1:16	Conform requirements
Growthtest	no growth	Conform

Date : 02 October 2023 Approved by QA/QC :



VERTROUWELIJK