



Doc. no. REA-S002-01-04-1606

Version

1.0

Date 02 Dec 2021 14:52:19 (GMT+01:00)

Page

1 of 2

Document owner

BBR

Title **CCC coombs control cells release**



Analysis	Requirement							Result
Antigen control	/	anti-A	anti-B	anti-A <sub>1</sub>	anti-D	DAT	IAT control	/
	CCC 3%	0	0	n.t.*	n.t.*	≥ 2+	1+ ≤ 2+	Conform requirements
Hematocrit	CCC 3 ± 1%							Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

\*n.t.: not tested.

Approved by

Date: 30 November 2023

QA/QC: 