



Doc. nr. REA-S008-03-03-0586

Versie 2.0

Datum 14 May 2021 10:23:38 (GMT+01:00)

Pagina 1 van 3

Document eigenaar BBR

Titel MASPAT kit release

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

K 1360

MASPAT kit

Monoclonal Antibody Solid Phase  
Platelet Antibody Test

IVD 96

REF K1360

LOT 8000460139

CE



2-8 °C

2025-04-11

(01)1872068203  
0025(10)800046  
0139(17)250411

Component	Requirement	Result
Positive control Lot: 8000460067	Positive reaction with: platelets Titration with: platelets $\geq$ 1:4	Conform requirements Conform requirements
Negative control Lot: 8000459782	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000459916	Titration with: platelets $\geq$ 1:4	Conform requirements
Microplates Lot: 8000460069	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000459462	Conductivity 3.4 – 4.8 mScm <sup>-1</sup> pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 07 December 2023 Approved by QA/QC :

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 28 Nov 2023 at: 08:01:29 am ; Printed by M.BOUWMAN