



Doc. nr. REA-S004-01-55-0566

Versie 2.0

Datum 06 May 2021 12:26:27 (GMT+02:00)

Pagina 1 van 2

Document eigenaar BBR

Titel Pelikloon anti-P₁ (IgM) monoclonal release

K1305

Clone P3NIL100

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Pelikloon
anti-P₁ (IgM) monoclonal v.
Clone: P3NIL100 2 ml

IVD 2-8°C REF K1305 CE
LOT 8000460228 2025-12-01

Analysis	Requirement	Results
Specificity	Negative with cells negative for the P ₁ antigen	Conform requirements
Avidity	Tube method	Conform requirements
	P ₁ ≥ 2+	
Titer	Tube method	Conform requirements
	P ₁ ≥ 1:4	
Growthtest	no growth	Conform

Date

: 27 February 2024

Approved by QA/QC

:

VERTROUWELIJK

Use copy according to REA-SOP-0002.
Printed on: 08 Jan 2024 at: 07:30:30 am ; Printed by A.ERASMUS