

Doc. nr. REA-S001-01-08-0487 Versie

2.0

Datum

06 May 2021 13:05:55 (GMT+02:00)

Pagina

Document eigenaar

1 van 2

**BBR** 

Titel

screeningspanel 123 release

K 1148



Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	3 ± 1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

: 21 March 2029 Approved by QA/QC