



Doc. nr. REA-S001-01-07-0486

Versie 2.0

Datum 06 May 2021 13:06:01 (GMT+02:00)

Pagina 1 van 3

Document eigenaar

BBR

Titel **Makropanel 16 (P) release**

K 1385

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Makropanel 16
3 ± 1%
16 x 3 mL

IVD
LOT 8000460429
REF K1385

www.sanquin.org/reagents
CE 0344
UDI

2-8 °C
 2024-05-02

(01)1872068203
1077(10)800046
0429(17)240502

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	3 ± 1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 21 March 2024 Approved by QA/QC :

VERTROUWELIJK