



Doc. nr. REA-S005-00-05-0577

Versie 2.0

Datum 14 May 2021 10:12:26 (GMT+01:00)

Pagina 1 van 2

Document eigenaar BBR

Titel PeliControl CcEeK+ release

K1399



Analysis	Requirement	Result
Control cells	CcDEeK+	Conform requirements
	DAT* negative	Conform requirements
Hematocrit	20 ± 2%	Conform requirements

\* = direct antiglobulin test

All the donor units and sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 28 March 2024 Approved by QA/QC

VERTROUWELIJK

Use copy according to REA-SOP-0002.  
Printed on: 18 Mar 2024 at: 03:08:10 pm ; Printed by M.DEMOL