


**Titel monospecifiek anti-human IgG release**

K1131




Analysis	Requirements	Result
Specificity	Negative with A, B and O DAT negative cells, tested in tube test method direct antiglobulin test (DAT) and in the indirect antiglobulin test (IAT) with PEG 4000 20%.	Conform requirements
Avidity/titer	Tested in Tube methode (DAT):  Results with IgG positive patient cells are at least comparable with reference.  Results with IgG positive in vitro sensitized cells are at least comparable with reference	Conform requirements  Conform requirements
Growth test	no growth	Conform

Date

: 12 April 2024

Approved by QA/QC

: 

VERTROUWELIJK