



Doc. nr. REA-S001-01-08-0487

Versie 2.0

Datum 06 May 2021 13:05:55 (GMT+02:00)

Pagina 1 van 2

Document eigenaar BBR

Titel screeningspanel 123 release

K 1148

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Screening panel 123
3 ± 1% 3 x 10 mL

CL
87110800460523
(17.240830)

IVD 2-8 °C
LOT 8000460523
REF K1148

www.sanquin.org 0344
reagents

2024-05-30 UDI

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	3 ± 1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

: 18 April 2024

Approved by QA/QC

:

VERTROUWELIJK

Use copy according to REA-SOP-0002.
Printed on: 09 Apr 2024 at: 02:19:57 pm ; Printed by M.DEMOL