



Doc. nr. REA-S004-01-58-0569

Versie 2.0

Datum 06 May 2021 12:26:29 (GMT+02:00)

Pagina 1 van 2

Document eigenaar BBR

Titel Pelikloon enhanced control release

K 1152



Analysis	Requirement	Result
Specificity	Negative with O, A ₁ and B cells	Conform requirements
Avidity	Detection of sensitized red cells as control for Jk ^b typing.	Conform requirements
Growthtest	No growth	Conform
External release	Released by Notified Body	Released

Date : 02 March 2021 Approved by QA/QC :

VERTROUWELIJK