

Doc. nr. REA-S001-01-09-0488 Versie

2.0

06 May 2021 13:47:24 (GMT+02:00) Datum

Pagina

Document eigenaar

1 van 2 **BBR**

Titel screeningsset 1 + 2 release

K 1388



Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	3 ± 1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

: 18 April 2024 Approved by QA/QC