



Doc. nr. REA-S001-01-03-0483

Versie 2.0


Datum 06 May 2021 13:05:51 (GMT+02:00)

Pagina 2 van 3

Document eigenaar BBR

Titel cellbind P3 (P) release

K7211



Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

: 19 April 2024

Approved by QA/QC

:

VERTROUWELIJK