



Title **Certificate of Analysis MAT Cell Set**

Product	MAT Cell Set
Product number	M2016

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

MAT Cell Set
 288

 ≤ -80°C

 2024-12-10

LOT 8000460721

REF M2016

LOT/ORG 68-7179

Label


Test results

Description	Specification	Meets specification
Qualification of donors and cells	Qualified according to European Pharmacopoeia chapter 2.6.30, section 5	yes
Donor infectious disease testing	All donors are screened and found negative for: <ul style="list-style-type: none"> Hepatitis B: HBsAg, HBV-NAT and anti-HBc Hepatitis C: Anti-HCV and HCV-NAT HIV: Anti-HIV1/2/(O) and HIV-NAT Treponema pallidum: anti-TP 	yes
Characteristics of cell pool	Consists of a pool of cryopreserved Peripheral Blood Mononuclear Cells (PBMC) isolated from four single donors	yes
	Cell concentration per vial after thawing from liquid nitrogen: <ul style="list-style-type: none"> 4 – 6 x 10⁶ cells/mL¹ 	yes
Monocyte Activation Test (MAT) ²	Detection of endotoxin: <ul style="list-style-type: none"> Mean OD_{450nm - 550nm} of the blank: <ul style="list-style-type: none"> < 0.1 Limit of Detection (based on mean blank+3x SD): <ul style="list-style-type: none"> ≤ 0.02 EU/mL³ Assurance of criteria for the endotoxin standard curve: <ul style="list-style-type: none"> Regression statistically significant: p < 0.01 Regression must not deviate significantly from linearity: p > 0.05 	yes
	Detection of non-endotoxin pyrogens (NEPs): <ul style="list-style-type: none"> Tested and found responsive towards NEPs: <ul style="list-style-type: none"> Peptidoglycan from <i>Staphylococcus aureus</i> Synthetic triacylated lipoprotein (Pam3CSK4) Flagellin from <i>Salmonella typhimurium</i> (FLA-ST) Heat-Killed <i>Staphylococcus aureus</i> (HKSA) 	yes
	Consideration of the averaging effect of pooling cells: <ul style="list-style-type: none"> EC₅₀ of endotoxin standard curve of each single donor: <ul style="list-style-type: none"> 50-200% of EC₅₀ of standard curve of the pool 	yes

¹ Determined by using an automated cell counter.

² Results obtained using PeliKine compact human IL-6 kit (REF: M1916) and PeliKine tool set 1 (REF: M1980).

³ Given as concentrations per sample (i.e. prior to addition of pMAT Cells to the samples).

Approved QA/QC	Date	17 June 2024	Signature	
----------------	------	--------------	-----------	---