



Doc. nr. REA-S004-01-46-0557

Versie 2.0

Datum 06 May 2021 11:56:11 (GMT+01:00)

Pagina 1 van 2

Document eigenaar BBR

Titel Pelikloon anti-Jka (IgM) monoclonal release

K 1300

Clone nr. MS15

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL
VL159/c
Pelikloon
anti-Jk^a (IgM) monoclonal
Clone: MS15
2mL
CE
0344
IVD 2-8°C REF K1300
LOT 8000460417 2026 07-06

Analysis	Requirement	Result
Specificity	Tube method Negative with cells negative for the Jk ^a antigen	Conform requirements
Avidity	Tube method	
	Jk (a+b-) ≥ 2+	Conform requirements
	Jk(a+b+) ≥ 2+	Conform requirements
Titer	Tube method	Conform requirements
	Jk(a+b+) ≥ 1:8	
Growthtest	no growth	Conform

Date

: 22 July 2024

Approved by QA/QC

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 11 Jul 2024 at. 09:52:46 am ; Printed by M.DEMOL