


Titel **Pelikloon anti-M (IgG) monoclonal release**

K 1302

Clone number LM110/140




Analysis	Requirement	Result
Specificity	Negative with cells negative for the M antigen	Conform requirements
Avidity	Tube Method	
	MM ≥ 3+	Conform requirements
	MN ≥ 2+	Conform requirements
Titer	Tube method	Conform requirements
	MN ≥ 1:4	
Growthtest	no growth	Conform

Date

: 09 July 2024

Approved by QA/QC

: 

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 13 Jun 2024 at: 07:32:08 am ; Printed by B SMIT