

**Titel MASPAT kit release**

K 1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

**MASPAT kit**  
**Monoclonal Antibody Solid Phase**  
**Platelet Antibody Test**

IVD  $\Sigma$  96

LOT 8000460774

REF K1360



2-8 °C

2026-01-10

(01)1872068203  
 0025(10)800046  
 0774(17)260110


Component	Requirement	Result
Positive control Lot: 8000460718	Positive reaction with: platelets Titration with: platelets $\geq$ 1:4	Conform requirements Conform requirements
Negative control Lot: 8000460310	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000460673	Titration with: platelets $\geq$ 1:4	Conform requirements
Microplates Lot: 8000460588	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000460308	Conductivity 3.4 – 4.8 mScm <sup>-1</sup> pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date

: 19 July 2024

Approved by QA/QC



VERTROUWELIJK