


**Titel Pelikloon anti-Jka (IgM) monoclonal release**

K 1300

Clone nr. MS15



Analysis	Requirement	Result
Specificity	Tube method	Conform requirements
	Negative with cells negative for the Jk <sup>a</sup> antigen	
Avidity	Tube method	
	Jk (a+b-) ≥ 2+	Conform requirements
	Jk(a+b+) ≥ 2+	Conform requirements
Titer	Tube method	Conform requirements
	Jk(a+b+) ≥ 1:8	
Growthtest	no growth	Conform

Date : 14 August 2024 Approved by QA/QC : 

VERTROUWELIJK