



Doc. nr. REA-S004-01-55-0566

Versie 2.0

Datum 06 May 2021 12:26:27 (GMT+02:00)


Pagina 1 van 2

Document eigenaar BBR

Titel **Pelikloon anti-P₁ (IgM) monoclonal release**

K1305

Clone P3NIL100



Analysis	Requirement	Results
Specificity	Negative with cells negative for the P ₁ antigen	Conform requirements
Avidity	Tube method	Conform requirements
	P ₁ ≥ 2+	
Titer	Tube method	Conform requirements
	P ₁ ≥ 1:4	
Growthtest	no growth	Conform

Date

: 01 August 2024

Approved by QA/QC

:

VERTROUWELIJK

Use copy according to REA-SOP-0002.
Printed on: 11 Jul 2024 at: 11:33:32 am ; Printed by B.SMIT