



Doc. nr. REA-S003-00-06-0500

Versie

2.0

Datum 06 May 2021 13:23:06 (GMT+01:00)

Pagina

1 van 2

Document eigenaar

BBR

Titel **Cellbind Screen release**

K 7000



Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

**Cellbind Screen**

IVD 2-8°C REF K7000



Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+	Conform requirements
	Results with anti- Fy <sup>a</sup> , anti- Fy <sup>b</sup> and anti-s are at least comparable with reference method	Conform requirements

Date

: 27 August 2024

Approved by QA/QC

:

VERTROUWELIJK

Use copy according to REA-SOP-0002.  
Printed on: 23 Jul 2024 at: 07:52:39 am ; Printed by T.VELDHON