



Doc. nr. REA-S004-01-42-0553

Versie 2.0

Datum 10 Jun 2021 09:23:34 (GMT+01:00)


Pagina 1 van 4

Document eigenaar BBR

Titel **Pelikloon anti-E (IgM) monoclonal release**

K1191

Clone number MS260



Analysis	Requirement	Result
Specificity	Negative with cells negative for the E antigen	Conform requirements
Avidity	Ee	
	Tube method $\geq 3+$	Conform requirements
	Microplate method $\geq 3+$	Conform requirements
	Cellbind Screen method $\geq 3+$	Conform requirements
	Cellbind Direct Type method $\geq 3+$	Conform requirements
Titer	Ee	
	Tube method $\geq 1:16$	Conform requirements
	Microplate method $\geq 1:16$	Conform requirements
Growthtest	no growth	Conform
External release	released by Notified Body	Released <input checked="" type="radio"/> Y <input type="radio"/> N

Date

: 24 October 2024

Approved by QA/QC

:

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 29 Aug 2024 at: 08:11:47 am ; Printed by B.SMIT