



Doc. nr. REA-S002-01-01-0493

Versie 2.0

Datum 06 May 2021 13:47:25 (GMT+02:00)

Pagina 1 van 2

Document eigenaar BBR

Titel Pelicontrol release

K1379



Analysis	Requirement	Result
Control cells	DAT* negative	Conform requirements
	sample 1: A ₂ B Rh D positive, negative screening.	Conform requirements
	sample 2: O Rh D negative, positive screening	Conform requirements
Hematocrit	20 ± 2%	Conform requirements

* = direct antiglobulin test

All the donor units and sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 05 December 2024 Approved by QA/QC

VERTROUWELIJK

Use copy according to REA-SOP-0002.
Printed on: 25 Nov 2024 at: 03:13:04 pm ; Printed by M.DEMOL