


Titel **Pelikloon anti-P1 (IgM) monoclonal release**

K1305

Clone P3NIL100



Analysis	Requirement	Results
Specificity	Negative with cells negative for the P ₁ antigen	Conform requirements
Avidity	Tube method	Conform requirements
	P ₁ ≥ 2+	
Titer	Tube method	Conform requirements
	P ₁ ≥ 1:4	
Growthtest	no growth	Conform

Date

: 22 May 2025

Approved by QA/QC

: 

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 21 May 2025 at: 10:17:55 am ; Printed by M.BOUWMAN