



**Titel Pelikloon anti-Lea (IgM) monoclonal release**

K 1306

Clone nr. LEA2



Analysis	Requirement	Result
Specificity	Negative with cells negative for the Le <sup>a</sup> antigen, tested in tube method	Conform requirements
Avidity	Tube method	Conform requirements
	Le(a+b-) ≥ 2+	
Titer	Tube method	Conform requirements
	Le(a+b -) ≥ 1:2	
Growthtest	no growth	Conform

Date : 06 May 2025 Approved by QA/QC : 

VERTROUWELIJK