



Doc. nr. REA-S003-00-06-0500

Versie 2.0


Datum 06 May 2021 13:23:06 (GMT+01:00)

Pagina 1 van 2

Document eigenaar BBR

Titel Cellbind Screen release

K 7000



Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Cellbind Screen


IVD 2-8°C REF K7000 6 CE

Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+	Conform requirements
	Results with anti- Fy ^a , anti- Fy ^b and anti-s are at least comparable with reference method	Conform requirements

Date

: 06 May 2025

Approved by QA/QC

: 

VERTROUWELIJK

Use copy according to REA-SOP-0002

Printed on 27 Mar 2025 at 02:50:33 pm, Printed by R.VANWIERINGEN