



Doc. Nr. REA-S-0114
Datum 04 Jul 2025

Revisie 03
Pagina 1 van 1
Document Eigenaar PROD-QC

Titel **MASPAT kit release**

K1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

MASPAT kit
Monoclonal Antibody Solid Phase
Platelet Antibody Test

  

   

8000461768
K1360

2-8 °C
2027-01-27

(01)1872068203
0025(10)800046
1768(17)270127

Component	Requirement	Result
Positive control Lot: 8000461401	Positive reaction with: platelets Titration with: platelets ≥ 1:4	Conform requirements Conform requirements
Negative control Lot: 8000461402	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000461340	Titration with: platelets ≥ 1:4	Conform requirements
Microplates Lot: 8000461716	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000461738	Conductivity 3.4 – 4.8 mScm ⁻¹ pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 24 July 2025 Approved by QA/QC : 