



Doc. No. REA-S-0028
Date 28 Apr 2025

Revision 03
Page 1 of 1
Document Owner PROD-QC

Title **Cellbind Screen release**

K7000



Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+	Conform requirements
	Results with anti- Fy ^a , anti- Fy ^b and anti-s are at least comparable with reference method	Conform requirements

Date

: 16 July 2025

Approved by QA/QC

:

CONFIDENTIAL