

Title **Cellbind Screen release**

K7000



Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Cellbind Screen


IVD    2-8°C REF K7000  

Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+	Conform requirements
	Results with anti- Fy ^a , anti- Fy ^b and anti-s are at least comparable with reference method	Conform requirements

Date

: 06 January 2026

Approved by QA/QC

