

Title Cellbind Screen release

K7000



Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Cellbind Screen

IVD  2-8°C  K7000  

Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+ Results with anti- Fy <sup>a</sup> , anti- Fy <sup>b</sup> and anti-s are at least comparable with reference method	Conform requirements Conform requirements

Date : 15 January 2026 Approved by QA/QC

