

Title **Cellbind Screen release**

K7000



Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Cellbind Screen

IVD **CE** **X** 2-8°C **REF** K7000 **W6** **€**

Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+ Results with anti- Fy ^a , anti- Fy ^b and anti-s are at least comparable with reference method	Conform requirements Conform requirements

Date : 15 January 2026 Approved by QA/QC



CONFIDENTIAL

Use copy according to REA-SOP-0002

Printed on 23 Dec 2025 at 08 00 28 am, Printed by R VANWIERINGEN