


Titel **Pelikloon anti-Jka (IgM) monoclonal release**

K1300
Clone nr. MS15



Analysis	Requirement	Result
Specificity	Tube method Negative with cells negative for the Jk ^a antigen	Conform requirements
Avidity	Tube method	
	Jk (a+b-) ≥ 2+ Jk(a+b+) ≥ 2+	Conform requirements Conform requirements
Titer	Tube method	
	Jk(a+b+) ≥ 1:8	Conform requirements
Growthtest	no growth	Conform

Date : 26 February 2026 Approved by QA/QC :

