

Titel **Pelicontrol release**

K1379



Analysis	Requirement	Result
Control cells	DAT* negative	Conform requirements
	sample 1: A ₂ B Rh D positive, negative screening.	Conform requirements
	sample 2: O Rh D negative, positive screening	Conform requirements
Hematocrit	20 ± 2%	Conform requirements

* = direct antiglobulin test

All the donor units and sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

19 February 2026

Approved by QA/QC



VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 10 Feb 2026 at 07:57:16 am ; Printed by M.DEMOL