



Doc. No. REA-S-0011  
Date 13 May 2025

Revision 03  
Page 1 of 2  
Document Owner PROD-QC

Title **cellbind P3 (P) release**


K7210

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL  
Cellbind P3  
0.5 ± 0.1%  
3 x 10 mL  
2-8 °C  
LOT 8000462190  
REF K7210  
2026-03-30  
UDI  
www.sanquin.org  
0344  
(01)187208820303  
8411018000462190  
(17)260330

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 13 February 2026 Approved by QA/QC

  
\_\_\_\_\_