



Doc. No. REA-S-0312

Revision 03

Date 04 Jul 2024

Page 1 of 1

Document Owner UVR

Title Certificate of Analysis MAT Cell Set

Product	MAT Cell Set	Label	Sanquin Plesmanlaan 125 1066 CX Amsterdam NL
Product number	M2016		MAT Cell Set ▽ 288 T \leq 80°C LOT 8000462184 2026 07 19 REF M2016 LOT ORG 68 7249

Test results

Description	Specification	Meets specification
Qualification of donors and cells	Qualified according to European Pharmacopoeia chapter 2.6.30, section 5	yes
Donor infectious disease testing	All donors are screened and found negative for. <ul style="list-style-type: none">• Hepatitis B: HBsAg, HBV-NAT and anti-HBc• Hepatitis C. Anti-HCV and HCV-NAT• HIV: Anti-HIV/1/2/(O) and HIV-NAT• <i>Treponema pallidum</i> anti-TP	yes
Characteristics of cell pool	Consists of a pool of cryopreserved Peripheral Blood Mononuclear Cells (PBMC) isolated from four single donors Cell concentration per vial after thawing from liquid nitrogen: <ul style="list-style-type: none">• $4 - 6 \times 10^6$ cells/mL¹	yes
Monocyte Activation Test (MAT) ²	<p>Detection of endotoxin:<ul style="list-style-type: none">• Mean OD_{450nm} – 550nm of the blank: < 0.1• Test sensitivity (0.02 EE/mL)³ > cut-off value• Assurance of criteria for the endotoxin standard curve:<ul style="list-style-type: none">◦ Coefficient of determination $R^2 > 0.975$◦ Regression statistically significant. $p < 0.01$◦ Lack of fit test: $p > 0.05$</p> <p>Detection of non-endotoxin pyrogens (NEPs)<ul style="list-style-type: none">• Tested and found responsive towards NEPs:<ul style="list-style-type: none">◦ Peptidoglycan from <i>Staphylococcus aureus</i>◦ Synthetic triacylated lipoprotein (Pam3CSK4)◦ Flagellin from <i>Salmonella typhimurium</i> (FLA-ST)◦ Heat-Killed <i>Staphylococcus aureus</i> (HKSA)</p> <p>Consideration of the averaging effect of pooling cells:<ul style="list-style-type: none">• EC₅₀ of endotoxin standard curve of each single donor<ul style="list-style-type: none">◦ 50-200% of EC₅₀ of standard curve of the pool</p>	yes

¹ Determined by using an automated cell counter ² Results obtained using PeliKine compact human IL-6 kit (REF M1916) and PeliKine tool set 1 (REF M1980) or PeliKine Human IL-6 ELISA Rapid Set (REFs M2018 and M2019)³.

³ Given as concentrations per sample (i.e. prior to addition of pMAT Cells to the samples)

Approved QA/QC	Date	05 February 2026	Signature	
----------------	------	------------------	-----------	--

CONFIDENTIAL

Use copy according to REA-SOP-0002
Printed on 02 Feb 2026 at 07 29.29 am, Printed by K TYMCHUK