

Title **Certificate of Analysis PeliKine Human IL-6 ELISA Rapid Set A (M2018)**

**Product information**

<b>Product name</b>	PeliKine Human IL-6 ELISA Rapid Set A	
<b>Product number</b>	M2018	
<b>Batch number</b>	8000462197	
<b>Label</b>	<p style="text-align: center; background-color: #4b4b4b; color: white; padding: 2px;">Sanquin Reagents B.V. Plesmanlaan 125 1066 CX Amsterdam NL</p> <h2 style="text-align: center;">PeliKine Human IL-6 ELISA Rapid Set A</h2> <p style="text-align: center;">Research Use Only Spin small volumes before use</p> <p style="text-align: center;">  -18 to -32°C              <span style="border: 1px solid black; padding: 2px;">LOT</span> 8000462197   2027-12-08              <span style="border: 1px solid black; padding: 2px;">REF</span> M2018         </p>	
	<p><b>USA</b> <b>Danger.</b> Causes severe skin burns and eye damage. May cause cancer.          Contains: sulfuric acid          Do not handle until all safety precautions have been read and understood. Do not breathe dusts or mists. Wear eye protection/face protection. If swallowed: Rinse mouth. Do NOT induce vomiting. If on skin (or hair): Take off immediately all contaminated clothing. Rinse skin with water/shower. If inhaled: Remove person to fresh air and keep comfortable for breathing. If in eyes: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a poison center/doctor. Specific treatment (see on this label). Wash contaminated clothing before reuse. Store locked up. Dispose of contents/container in accordance with local/regional/national/international regulations.</p> <p><b>EU</b> Safety data sheet available on request.          Tel: +31 20 512 3599 Mon-Fri 9:00-17:00 CET</p>	

**Test results**

Description	Specification	Meets specification
Production and release of product	According to Essange Reagents' certificated ISO13485 quality management system	yes
ELISA	IL-6 standard curve: - Tested concentrations: blank, 0.6, 1.9, 5.6, 16.7, 50, 150 and 450 pg/mL - Sensitivity (based on mean of blank + 3x standard deviation): ≤ 1 pg/mL - Mean OD450-550nm of the blank: ≤ 0.140	yes

Date: 10 February 2026

Approved by QA: 