



Doc. No. REA-S-0023
Date 30 May 2025

Revision 05
Page 1 of 1
Document Owner PROD-QC

Title **MASPAT Indicator Red Cells release**

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL
MASPAT Indicator Red Cells ^{VL}
0.3 ± 0.1%
10 mL
2-8 °C
Lot 8000462341
REF K1139 2026 05 31 UDI

Analysis	Requirement				Result
		anti-A	anti-B	DAT	
Antigen control	MIRC 0.3%	0	0	≥ 2+	Conform requirements
Hematocrit	MIRC 0.3 ± 0.1%				Conform requirements

The donor unit used in the production of this reagent has been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date: 16 APRIL 2026

Approved by
QA: 