

Title **cellbind P3 (P) release**

K7210

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Cellbind P3
0.5 ± 0.1%
3 x 10 ml

IVD 2-8 °C
LOT 8000462318
REF K7210

www.sanquin.org
reagents 0344

2026 05 26

UDI

01118720620303
841108000462318
1171280526

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 13 April 2026 Approved by QA/QC : 