



Doc. No. REA-S-0009
Date 28 Apr 2025

Revision 03
Page 1 of 2
Document Owner PROD-QC

Title **cellbind ID 16 (P) release**

K7230

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL
Cellbind ID16^{CL}
0.5 ± 0.1%
16 x 3 mL

www.sanquin.org/reagents

CE 0344 UDI

IVD
LOT 8000462319
REF K7230

2-8 °C
2026-05-26

(01)1872068203
1107(10)800046
2319(17)260526

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

: 13 April 2026

Approved by QA/QC

:

CONFIDENTIAL

Use copy according to REA-SOP-0002.

Printed on: 30 Mar 2026 at: 02:50:47 pm, Printed by M DEMOL