



Doc. No. REA-S-0023  
Date 30 May 2025

Revision 05  
Page 1 of 1  
Document Owner PROD-QC

Title MASPAT Indicator Red Cells release

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL  
MASPAT Indicator Red Cells<sup>VL</sup>  
0.3 ± 0.1%  
10 mL  
IVD 2-8 °C  
LOT 8000462478  
REF K1139 2026-07-26 UDI

Analysis	Requirement			Result
Antigen control		anti-A	anti-B	DAT
	MIRC 0.3%	0	0	≥ 2+
Hematocrit	MIRC 0.3 ± 0.1%			Conform requirements

The donor unit used in the production of this reagent has been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date: 11 June 2026

Approved by

QA: 

CONFIDENTIAL