

Titel **Makropanel 16 (P) release**

K1385

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Makropanel 16

3 ± 1%

16 x 3 mL

IVD

LOT 8000462473

REF K1385



UDI

0344



2-8 °C

2026-07-20

(01)1872068203

1077(10)800046

2473(17)260720

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigenogram of the panel	Conform requirements
Hematocrit	3 ± 1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

: 05 June 2026

Approved by QA/QC

