



Doc. No. REA-S-0009  
Date 28 Apr 2025

Revision 03  
Page 1 of 2  
Document Owner PROD-QC

Title cellbind ID 16 (P) release

K7230

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL  
Cellbind ID16  
0.5 ± 0.1%  
16 x 3 mL<sup>CL</sup>  
www.sanquin.org/reagents  
0344  
UDI  
IVD  
LOT 8000462476  
REF K7230  
2-8 °C  
2026-07-20  
(01)1872068203  
1107(10)800046  
2476(17)260720

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigenogram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

: 05 June 2026

Approved by QA/QC

:

CONFIDENTIAL