



Doc. No. 3030
Date 17/06/2026

Version 4.0
Page 1 of 2
Document Owner PROD-BGS

Title **Makropanel 16 (P) release**

K1385

 **Sanquin Plesmanlaan 125-1066 CX Amsterdam NL**

Makropanel 16
 $3 \pm 1\%$

16 x 3 mL^{CL}

 **IVD**
 **LOT** 8100000017
 **REF** K1385

 www.sanquin.org/reagents
 **CE** 0344
 **UDI**

 2-8 °C
2026-08-17


(01)1872068203
1077(10)810000
0017(17)260817

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	$3 \pm 1\%$	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 08 July 2026 Approved by QA/QC

: 

CONFIDENTIAL