



Doc. No. REA-S-0028

Date 28 Apr 2025

Revision 03

Page 1 of 1

Document Owner PROD-QC

Title **Cellbind Screen release**

K7000



Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Cellbind Screen

IVD 2-8°C REF K7000 6

Analysis	Requirement	Result
Specificity	Negative reaction with A, B and O cells	Conform requirements
	Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women	Conform requirements
Avidity / Titer	Reaction with Coombs Control Cells tested with AB serum is at least 3+	Conform requirements
	Results with anti- Fy ^a , anti- Fy ^b and anti-s are at least comparable with reference method	Conform requirements

Date

: 31 July 2026

Approved by QA/QC

:

CONFIDENTIAL

Use copy according to Doc. ID4016, Printed on: 20/07/2026